

Certificate of Compliance



No. 2J190819.TFSQQ92

Technical Construction File no. B1908TR8129S, B1908TR8129E

Certificate's
Holder:

Thermo Fisher Scientific (China)
Co., Ltd.

Building 6, No.27, Xin Jinqiao Road, Pudong,
Shanghai, China

Certification ECM
Mark:



Product:
Model(s):

Rock/Rotator Series
88882001, 88882002, 88882003, 88882004,
88882xxx

Verification to:

Standard:
EN 61010-1:2010, EN 55011:2016+A.1:2017,
EN 61326-1:2013, EN 61000-3-2:2014,
EN 61000-3-3:2013

related to CE Directive(s):
2014/35/EU (Low Voltage)
2014/30/EU (Electromagnetic Compatibility)

Remark: The product(s) has been verified on a voluntary basis. The product(s) satisfies the requirements of the Certification Mark of ECM, in reference to the above listed Standard(s). The above Compliance Mark can be affixed on the product(s) accordingly to the ECM regulation about its release and its use. The regulation can be found at www.entecerma.it. This Certificate of Compliance can be checked for validity at www.entecerma.it.

This verification doesn't imply assessment of the production of the product(s).

Additional information, clarification about the **CE** Marking:



We attest that a TCF for the **CE** Marking process is in place. Whereas the Manufacturer is Responsible to start the **CE Marking Certification Procedure** and to perform all the necessary activities, as required by the Directive before placing the **CE** Mark on the product(s).

Date of issue 19 August 2019

Expiry date 18 August 2024

Chief Manager
Marco Morino

Deputy Manager
Amanda Payne

Ente Certificazione Macchine Srl

Via Ca' Bella, 243 – Loc. Castello di Serravalle – 40053 Valsamoggia (BO) - ITALY
☎ +39 051 6705141 📠 +39 051 6705156 ✉ info@entecerma.it 🌐 www.entecerma.it