



Wir / We / Nous:

Phoenix Instrument GmbH

erklären in alleiniger Verantwortung, dass folgendes Produkt / declare under our sole responsibility that the following product / déclarons sous notre propre responsabilité que le produit ci-dessous

Labor-Magnetrührer / Laboratory magnetic stirrer:

Modell / Model / Modèle:

RSM-03-4KH

auf das sich diese Erklärung bezieht, mit den folgenden Normen oder normativen Dokumenten übereinstimmt: / to which this declaration relates, is in conformity with the following standards or other normative documents: / auquel se réfère cette déclaration est conforme aux normes ou autres documents normatifs:

EN 61010-1	Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use – Part 1: General Requirements
EN 61010-2-010:2014	Part 2-010: Particular requirements for laboratory equipment for the heating of materials
EN 61010-2-051:2003	Part 2-051: Particular requirements for laboratory equipment for mixing and stirring
EN 61326-1	Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
EN 61000-3-2	Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions
EN 61000-3-3	Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker

sowie den Bestimmungen der folgenden Richtlinien entspricht: / following the provisions of directives: / conformément aux dispositions des directives:

2014/35/EU	Low Voltage Directive
2014/30/EU	Electromagnetic Compatibility Directive
2015/863/EU	RoHS Directive

Garbsen, den 30.04.2021



Christian Lieske
Managing Director



Marco Riemer
Managing Director

Phoenix Instrument GmbH
Heinkelstraße 4
30827 Garbsen
Tel. 05131/90818-30
www.info@phoenix-instrument.de
info@phoenix-instrument.de