



Wir / We / Nous:

Phoenix Instrument GmbH

erklären in alleiniger Verantwortung, dass folgendes Produkt / declare under our sole responsibility that the following product / déclarons sous notre propre responsabilité que le produit ci-dessous

Zentrifuge / Centrifuge:

Modell / Model / Modèle:

CD-0424

auf das sich diese Erklärung bezieht, mit den folgenden Normen oder normativen Dokumenten übereinstimmt: / to which this declaration relates, is in conformity with the following standards or other normative documents: / auquel se réfère cette déclaration est conforme aux normes ou autres documents normatifs:

- EN 61010-1 **Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use – Part 1: General Requirements**
- EN 61326-1 **Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements**
- EN 61000-3-2 **Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions**
- EN 61000-3-3 **Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker**
- EN 61010-2-020 **Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 2-020: Particular requirements for laboratory centrifuges**

sowie den Bestimmungen der folgenden Richtlinien entspricht: / following the provisions of directives: /conformément aux dispositions des directives:

- 2014/35/EU **Low Voltage Directive**
- 2014/30/EU **Electromagnetic Compatibility Directive**
- 2015/863/EU **RoHS Directive**

Garbsen, den 03.09.2022



Christian Lieske
Managing Director

Marco Riemer
Managing Director

Phoenix Instrument GmbH
Heinkelstraße 4
30827 Garbsen
Tel. 05131/90818-30
www.info@phoenix-instrument.de
info@phoenix-instrument.de