

EC DECLARATION OF CONFORMITY

Cleaver Scientific Ltd. declare under their sole responsibility that the following units comply and are in conformity with the statutory CE safety directives:

Product Name: gelLITE Gel Documentation System

Product Codes: Beginning GELLITE

These products conform to the following directives:

- EMC Directive 2014/30/EU
- Low Voltage Directive 2014/35/EU
- RoHS Directive 2011/65/EU

and carry the CE mark accordingly.

The above conformity is based on compliance of the product with the following harmonised standards:

EMC Directive as per: 61326-2-6:2013 Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements.

According to the following standards:

EN 61000-4-2	Electrostatic discharge (ESD)
EN 61000-4-3	Radiated RF immunity
EN 61000-4-4	Electrical fast transient/burst
EN 61000-4-5	Surge immunity
EN 61000-4-6	Conducted RF immunity
EN 61000-4-11	Voltage dips and interruptions
EN 55011, Class A	RF emissions
EN 61000-3-2	Limits for harmonic current emissions
EN 61000-3-3	Voltage fluctuations and flicker

Low Voltage Directive as per: EN 61010-1:2010

Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use


For and on behalf of:

Cleaver Scientific Ltd



Mr Adrian Cleaver

Managing Director



Certificate No. 17533