



Wir / We / Nous:

Phoenix Instrument GmbH

erklären in alleiniger Verantwortung, dass folgendes Produkt / declare under our sole responsibility that the following product / déclarons sous notre propre responsabilité que le produit ci-dessous

Laborschüttler / Laboratory Shaker

Modell / Model / Modèle:

RS-OS 5, RS-LS 5, RS-RS 5, RS-DS 5

auf das sich diese Erklärung bezieht, mit den folgenden Normen oder normativen Dokumenten übereinstimmt: / to which this declaration relates, is in conformity with the following standards or other normative documents: / auquel se réfère cette déclaration est conforme aux normes ou autres documents normatifs:

- EN 61010-1 Safety Requirements for Electrical Equipment for Measurement Control, and Laboratory Use – Part 1: General Requirements
- EN 61326-1 Electrical equipment for measurement, control and laboratory use - EMC requirements - Part 1: General requirements
- EN 61000-3-2 Electromagnetic compatibility (EMC) - Part 3-2: Limits - Limits for harmonic current emissions
- EN 61000-3-3 Electromagnetic compatibility (EMC) - Part 3-3: Limits - Limitation of voltage changes, voltage fluctuations and flicker

sowie den Bestimmungen der folgenden Richtlinien entspricht: / following the provisions of directives: /conformément aux dispositions des directives:

- 2014/35/EU Low Voltage Directive
- 2014/30/EU Electromagnetic Compatibility Directive
- 2015/863/EU RoHS Directive

Garbsen, den 13.01.2020



Christian Lieske
Managing Director

Marco Riemer
Managing Director

Phoenix Instrument GmbH
Heinkelstraße 4
30827 Garbsen
Tel. 05131/90818-30
www.info@phoenix-instrument.de
info@phoenix-instrument.de