

Specification

Item number: 2432

Conductivity Standard

ROTI@Calipure 84 $\mu\text{S}/\text{cm}$ (25 °C)

CAS No.:

print date:

01.01.2024

formula:

density: 1

molecular weight:

Type analysis

Conductivity (25°C)

84 $\mu\text{S}/\text{cm} \pm 1 \%$

Traceable to SRM from NIST

The following values of the temperature table are not measured in each batch. Values are only informative and not content of our product specification.

Temperature	Conductivity value ($\mu\text{S}/\text{cm}$)
5°C	53,02
10°C	60,34
15°C	67,61
16°C	69,25
17°C	70,89
18°C	72,52
19°C	74,16
20°C	75,8
21°C	77,44
22°C	79,08
23°C	80,72
24°C	82,36
25°C	84
30°C	92,19
35°C	100,92
40°C	109,21
45°C	118,05
50°C	126,8

Our products are tested for laboratory use only.

The stated information refers to the current status of product quality.

We reserve the right to implement necessary changes.

Dr. R. Niemand

Head of Quality Assurance

S. Lorsee

Head of Quality Management

This document is computer printout and has therefore not been signed by hand.

Carl Roth GmbH + Co. KG
Schoemperlenstraße 3-5
76185 Karlsruhe

Telefon 0721/5606-0
Telefax 0721/5606-149
E-Mail: info@carlroth.de

Die Firma ist eine Kommanditgesellschaft mit Sitz in Karlsruhe, Reg. Gericht Mannheim HRA 100055. Persönlich haftende Gesellschafterin ist die Firma Roth Chemie GmbH mit Sitz in Karlsruhe, Reg. Gericht Mannheim HRB 100428. Geschäftsführer: André Houdelet

