

Specification

Item number:	9648		
Roti®-Aquatest Plate Strep			
DEV, ISO 7899-2, ready-to-use, sterile, for microbiology			
CAS No.:		print date:	01.02.2019
formula:			
density:		storage temperature:	+4 °C
molcular weight:			

Composition

di-Potassium hydrogen phosphate	4 g/l
Sodium azide	0,4 g/l
Tetrazolium chloride (TTC)	0,1 g/l
Tryptose	20 g/l
Yeast extract	5 g/l
Glucose	2 g/l
Agar	12 g/l
pH-value	7,2 ±0,1
Colour	amber, slightly opalescent
Growth of test strains	complies
Diameter	55 mm

Enumeration of faecal streptococci

Our products are tested for laboratory use only.
The stated information refers to the current status of product quality.
We reserve the right to implement necessary changes.

U. Jeckel
Quality Assurance

S.Hildebrandt
Quality Management

This document is computer printout and has therefore not been signed by hand.

Carl Roth GmbH + Co. KG
Schoemperlenstraße 3-5
76185 Karlsruhe

Telefon 0721/5606-0
Telefax 0721/5606-149
E-Mail: info@carlroth.de

Die Firma ist eine Kommanditgesellschaft mit Sitz in Karlsruhe, Reg. Gericht Mannheim HRA 100055. Persönlich haftende Gesellschafterin ist die Firma Roth Chemie GmbH mit Sitz in Karlsruhe, Reg. Gericht Mannheim HRB 100428. Vorsitzender des Aufsichtsrats: Eberhard Gaul
Geschäftsführer: Lothar Haidmann

